

Recommended protocol for IgK

Obtained in run

21 Aug 2016

Immunostainer

Type: Dako Autostainer Link 48 +

Primary antibody

Clone: Polyclonal
Producer: Dako
Product no. / lot no.: A0191 / 040(301)
Diluent: Antibody Diluent
Diluent producer / no.: Dako / S2022
Dilution factor: 1:3000
Incubation time / temperature: 60 min. / 20°C

Epitope retrieval, HIER

Device: Microwave oven
Buffer: Std Citrate pH 6
Heating time at max. temp.: 15 min.
Maximum heating temp.: 100°C

Visualization system

Producer: Dako
Product / no: EnVision FLEX / K8000/SM802
Linker: None
Incubation time polymer: 30 min.
Incubation temperature: 20°C

Chromogen

Producer: Dako
Product / no: DAB+ / DM827 / DM823 / K800
Incubation time / temperature: 10 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.