

Recommended protocol for WT1

Obtained in run 55

10 Dec 2018

Immunostainer

Type: Thermo Autostainer 36/48/72

Primary antibody

Clone: WT49
Producer: Leica/Novocastra
Product no. / lot no.: NCL-L-WT1-562 / 6062869
Diluent: normal antibody diluent
Diluent producer / no.: Immunologic / VWRKBD09-999
Dilution factor: 1:10
Incubation time / temperature: 60 min. / 20°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer: Dako TRS High pH (3-1)
Heating time at max. temp.: 10 min.
Maximum heating temp.: 97°C

Visualization system

Producer: Dako
Product / no: EnVision FLEX+ / K8002/SM802
Linker: None
Incubation time polymer: 30 min.
Incubation temperature: 20°C

Chromogen

Producer: Dako
Product / no: DAB+ / DM827 / DM823 / K800
Incubation time / temperature: 10 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.