

Recommended protocol for CD15

Obtained in run 61

22 Dec 2020

Immunostainer

Type: Dako Autostainer

Primary antibody

Clone: Carb-3
Producer: Dako
Product no. / lot no.: M3631 / 10109709
Diluent: Antibody Diluent
Diluent producer / no.: Dako / S0809
Dilution factor: 1:50
Incubation time / temperature: 60 min. / 22°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer: Dako TRS High pH (3-1)
Heating time at max. temp.: 20 min.
Maximum heating temp.: 97°C

Visualization system

Producer: Dako
Product / no: EnVision FLEX / K8000/SM802
Linker: None
Incubation time polymer: 30 min.
Incubation temperature: 22°C

Chromogen

Producer: Dako
Product / no: DAB+ / DM827 / DM823 / K800
Incubation time / temperature: 10 min. / 22°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.