

Recommended protocol for CK7

Obtained in run 62

13 Apr 2021

Immunostainer

Type: Thermo Autostainer 36/48/72

Primary antibody

Clone: OV-TL 12/30
Producer: ThermoScientific
Product no. / lot no.: MS-1352-P / 1352P1912B
Diluent: Normal Antibody Diluent
Diluent producer / no.: Immunologic / VWRKBD09-999
Dilution factor: 1:200
Incubation time / temperature: 60 min. / 25°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer: Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.: 20 min.
Maximum heating temp.: 99°C

Visualization system

Producer: Immunologic
Product / no: Mouse/Rabbit HRP / VWRKDPVB999HRP
Linker: None
Incubation time polymer: 30 min.
Incubation temperature: 25°C

Chromogen

Producer: Scytek
Product / no: Dab chromogen / ACB060
Incubation time / temperature: 8 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.