

Recommended protocol for CD138

Obtained in run 69

27 Jul 2023

Immunostainer

Type: Dartmon - AS330PLUS

Primary antibody

Clone: DA013
Producer: Shenzhen Dartmon Biotechnology
Product no. / lot no.: RMB1A054 / 0542306191
Diluent: None
Diluent producer / no.: None / None
Dilution factor: 1:0
Incubation time / temperature: 30 min. / 25°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Dartmon Immunohistochemical Antigen Retrieval Buffer
Heating time at max. temp.: 20 min.
Maximum heating temp.: 100°C

Visualization system

Producer: Dartmon
Product / no: Immunochromogenic Reagent / DMRD4044
Linker: Immunochromogenic Reagent
Incubation time linker: 20 min.
Incubation time polymer: 20 min.
Incubation temperature: 25°C

Chromogen

Producer: Dartmon
Product / no: Immunochromogenic Reagent / DMRD4044
Incubation time / temperature: 6 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.