

Recommended protocol for MSH6

Obtained in run 72

25 Jul 2024

Immunostainer

Type: Ventana Benchmark Ultra

Primary antibody

Clone: EPR3945
Producer: Abcam
Product no. / lot no.: ab92471 / GR3368591-8
Diluent: Antibody Dilution Buffer
Diluent producer / no.: Ventana / ABD250
Dilution factor: 1:1000
Incubation time / temperature: 48 min. / 36°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Ventana Ultra CC1
Heating time at max. temp.: 64 min.
Maximum heating temp.: 100°C

Visualization system

Producer: Ventana
Product / no: OptiView DAB IHC Detection Kit / 760-700
Incubation time linker: 4 min.
Incubation time polymer: 4 min.
Amplifier: OptiView Amplification Kit
Incubation time amplifier: 4+4 min.
Incubation temperature: 36°C

Chromogen

Producer: Ventana
Product / no: OptiView DAB IHC Detection Kit / 760-700
Incubation time / temperature: 4 min. / 36°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect



Nordic Immunohistochemical Quality Control

Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

the most recent developments in immunohistochemistry at any point in time.