

Recommended protocol for PD-L1 (lung)

Obtained in run C2

25 Aug 2017

Immunostainer

Type: Ventana Benchmark XT

Primary antibody

Clone: SP263
Producer: Ventana
Product no. / lot no.: 790-4905 / G06399Z
Format: Ready-To-Use (prediluted)
Incubation time / temperature: 16 min. / 37°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Ventana CC1
Heating time at max. temp.: 64 min.
Maximum heating temp.: 95°C

Visualization system

Producer: Ventana
Product / no: OptiView DAB IHC Detection Kit / 760-700
Incubation time linker: 12 min.
Incubation time polymer: 12 min.
Amplifier: OptiView Amplification Kit
Incubation time amplifier: 8 min.
Incubation temperature: 37°C

Chromogen

Producer: Ventana
Product / no: OptiView DAB IHC Detection Kit / 760-700
Incubation time / temperature: 8 min. / 37°C
Enhancement: CuSO₄

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.